

湖南臻和亦康医疗用品有限公司

HUNAN ZHENHEYIKANG MEDICAL INSTRUMENT CO, LTD

检验报告单

INSPECTION REPORT

编号/NO.: JL-ZG-CG025 A/0

产品名称 Product name	医用棉球 Medical cotton ball	生产批号 Lot No	20230323
型号规格 Specification	0.4g/grain 25g/bag	灭菌批号 Sterilization Lot No	/
检验日期 Inspection Date	2023.03.24	批 量 Sampling Quantity	/
检测项目 Test item	标准要求 Standard	检验结果 Inspection result	
物 理 physics	3.1 Weight requirement	0.4g±0.04g 25g±2.5g	Qualified
	3.2.1 外观要求 Appearance requirements	棉球的棉花纤维应柔软、白色、无臭，不得有黄斑、污渍、异物。 The cotton fiber of the cotton ball should be soft, white, odorless, no macular, stains, foreign matter.	Qualified
	3.2.2 吸水时间 Water absorption time	棉球应于 10 秒内吸水沉入液面以下。The cotton ball should absorb water and sink below the liquid level within 10 seconds.	Qualified
	3.2.3 吸水量 soakage	每克棉球的吸水量应不少于 23g。The water absorption per gram of cotton ball should not be less than 23g.	Qualified
	3.2.4 包装外观 Packaging appearance	包装完整、无破损。外包装上的标签填写应清楚、整齐，内容应包括：制造商名、产品名称、生产批号等。The package is complete and without breakage. The label on the outer package shall be filled in clearly and neatly, which shall include: manufacturer's name, product name, production batch number, etc production batch number, etc.	Qualified

结 论 conclus ion	<p>按湘潭械备 20150010 号医用棉球技术要求进行检测，结果符合标准的要求。</p> <p>The test was carried out according to the technical requirements of Xiangtan Equipment 20150010, and the results met the requirements of the standard.</p> <p>(质管部盖章)</p> <p>签发日期 Date of issue : 2023 年 03 月 24 日</p>
-------------------------------	--

Inspector:

Reviewer:

Check: